

Analytical Results For:

 STORM CONSTRUCTION LLC
 STEPHEN WILSON
 P. O. BOX 1552
 HOPE NM, 88250
 Fax To: (575) 748-8841

 Received: 08/12/2016
 Reported: 08/18/2016
 Project Name: BRIGHAM 'B'
 Project Number: NONE GIVEN
 Project Location: SEC 22-T17S-R30E

 Sampling Date: 08/12/2016
 Sampling Type: Soil
 Sampling Condition: ** (See Notes)
 Sample Received By: Jodi Henson

Sample ID: #3 (H601795-03)

TPH 8015M		mg/kg		Analyzed By: MS					
Analyte	Result	Reporting Limit	Analyzed	Method Blank	BS	% Recovery	True Value QC	RPD	Qualifier
GRO C6-C10	<10.0	10.0	08/12/2016	ND	197	98.4	200	0.494	
DRO >C10-C28	<10.0	10.0	08/12/2016	ND	197	98.7	200	0.0887	
Surrogate: 1-Chlorooctane	86.5 %	35-147							
Surrogate: 1-Chlorooctadecane	97.2 %	28-171							


Sample ID: #4 (H601795-04)

TPH 8015M		mg/kg		Analyzed By: MS					
Analyte	Result	Reporting Limit	Analyzed	Method Blank	BS	% Recovery	True Value QC	RPD	Qualifier
GRO C6-C10	<10.0	10.0	08/12/2016	ND	197	98.4	200	0.494	
DRO >C10-C28	<10.0	10.0	08/12/2016	ND	197	98.7	200	0.0887	
Surrogate: 1-Chlorooctane	85.6 %	35-147							
Surrogate: 1-Chlorooctadecane	96.6 %	28-171							

Cardinal Laboratories

* = Accredited Analyte

PLEASE NOTE: Liability and Damages. Cardinal's liability and client's exclusive remedy for any claim arising, whether based in contract or tort, shall be limited to the amount paid by client for analyses. All claims, including those for negligence and any other cause whatsoever shall be deemed waived unless made in writing and received by Cardinal within thirty (30) days after completion of the applicable service. In no event shall Cardinal be liable for incidental or consequential damages, including, without limitation, business interruptions, loss of use, or loss of profits incurred by client, its subsidiaries, affiliates or successors arising out of or related to the performance of the services hereunder by Cardinal, regardless of whether such claim is based upon any of the above stated reasons or otherwise. Results relate only to the samples identified above. This report shall not be reproduced except in full with written approval of Cardinal Laboratories.



Celey D. Keene, Lab Director/Quality Manager

Analytical Results For:

 STORM CONSTRUCTION LLC
 STEPHEN WILSON
 P. O. BOX 1552
 HOPE NM, 88250
 Fax To: (575) 748-8841

 Received: 08/12/2016
 Reported: 08/18/2016
 Project Name: BRIGHAM 'B'
 Project Number: NONE GIVEN
 Project Location: SEC 22-T17S-R30E

 Sampling Date: 08/12/2016
 Sampling Type: Soil
 Sampling Condition: ** (See Notes)
 Sample Received By: Jodi Henson

Sample ID: #1 (H601795-01)

TPH 8015M		mg/kg	Analyzed By: MS						
Analyte	Result	Reporting Limit	Analyzed	Method Blank	BS	% Recovery	True Value QC	RPD	Qualifier
GRO C6-C10	<10.0	10.0	08/12/2016	ND	197	98.4	200	0.494	
DRO >C10-C28	<10.0	10.0	08/12/2016	ND	197	98.7	200	0.0887	
Surrogate: 1-Chlorooctane	91.4 %	35-147							
Surrogate: 1-Chlorooctadecane	100 %	28-171							

Sample ID: #2 (H601795-02)

TPH 8015M		mg/kg	Analyzed By: MS						
Analyte	Result	Reporting Limit	Analyzed	Method Blank	BS	% Recovery	True Value QC	RPD	Qualifier
GRO C6-C10	<10.0	10.0	08/12/2016	ND	197	98.4	200	0.494	
DRO >C10-C28	<10.0	10.0	08/12/2016	ND	197	98.7	200	0.0887	
Surrogate: 1-Chlorooctane	88.6 %	35-147							
Surrogate: 1-Chlorooctadecane	93.0 %	28-171							

Cardinal Laboratories

* = Accredited Analyte

PLEASE NOTE: Liability and Damages. Cardinal's liability and client's exclusive remedy for any claim arising, whether based in contract or tort, shall be limited to the amount paid by client for analyses. All claims, including those for negligence and any other cause whatsoever shall be deemed waived unless made in writing and received by Cardinal within thirty (30) days after completion of the applicable service. In no event shall Cardinal be liable for incidental or consequential damages, including, without limitation, business interruptions, loss of use, or loss of profit incurred by client, its subsidiaries, affiliates or successors arising out of or related to the performance of the services hereunder by Cardinal, regardless of whether such claim is based upon any of the above stated reasons or otherwise. Results relate only to the samples identified above. This report shall not be reproduced except in full with written approval of Cardinal Laboratories.



Celey D. Keene, Lab Director/Quality Manager

August 18, 2016

STEPHEN WILSON

STORM CONSTRUCTION LLC

P. O. BOX 1552

HOPE, NM 88250

RE: BRIGHAM 'B'

Enclosed are the results of analyses for samples received by the laboratory on 08/12/16 9:30.

Cardinal Laboratories is accredited through Texas NELAP under certificate number T104704398-16-8. Accreditation applies to drinking water, non-potable water and solid and chemical materials. All accredited analytes are denoted by an asterisk (*). For a complete list of accredited analytes and matrices visit the TCEQ website at www.tceq.texas.gov/field/ga/lab_accred_certif.html.

Cardinal Laboratories is accredited through the State of Colorado Department of Public Health and Environment for:

Method EPA 552.2	Halooacetic Acids (HAA-5)
Method EPA 524.2	Total Trihalomethanes (TTHM)
Method EPA 524.4	Regulated VOCs (V1, V2, V3)

Accreditation applies to public drinking water matrices.

This report meets NELAP requirements and is made up of a cover page, analytical results, and a copy of the original chain-of-custody. If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Caley D. Keene

Lab Director/Quality Manager