

Exhibit II

Independent Chloride Lab Results of SM4500 Cl B vs. EPA 300

| Sample | Chloride mg/kg via EPA300 | Chloride mg/kg via SM4500 Cl B | RPD |
|--------|---------------------------|--------------------------------|-------|
| 1 | 189 | 200 | -5.66 |
| 2 | 426 | 440 | -3.23 |
| 3 | 4540 | 4500 | 0.88 |
| 4 | 2 | <10 | 0.00 |
| 5 | 1490 | 1580 | -5.86 |

*Sample results were performed by Alamo Analytical Laboratories, an independent laboratory. Alamo Analytical is a NELAP certified laboratory through the State of Texas. Analyses were performed on a 1:10 DI water extraction. Chart representation titled "Independent Lab Study Data of EPA 300 vs. SM4500 Cl B."

Independent Lab Study Data of EPA 300 vs. SM4500 Cl B

