


BW-22


**Wasserhund
Tatum**

**MIT
9/9/13**



D & L Meters & Instrument Service, Inc.

Lovington, NM 88260
P.O. Box 1621
Office: (575) 396-3715
Fax: (575) 396-5812




Friday, June 28, 2013

Certification of Pressure Recorder Test:

Company: Celero Energy
Unit: N/A
Model: 8" Bristols
Pressure Rating: 1,000#
Serial #: 3914

*This Pressure Recorder was tested at midrange for accuracy and verified within
+5% and -5% for a 1,000# pressure element.*


Jesse Arenivas, Technician

BLOW-OUT PREVENTERS
FLUID CONTAINMENT SYSTEMS