


10.222

D & L Meters & Instrument Service, Inc.

,0

Lovington, NM 88260 P.O. Box 1621 Office: (575) 396-3715 (575) 396-5812 Fax:



CONTAIN

Invoice # 100177

Certification of Pressure Recorder Test:

Company: Gandy Unit: 2 Model: 8"Chartrecorder **Pressure Rating:** 1,000#

Serial #:

This Pressure Recorder was tested at midrange for accuracy and verified within +5% and -5% for 1,000# pressure element.

Issac Luna