

Exhibit III

Low Level Chloride Comparison Data for EPA 300 vs. SM4500 Cl B

Sample	Chloride mg/kg via EPA300	Chloride mg/kg via SM4500	RPD
1	76.7	128	-50.12
2	26.7	64	-82.25
3	27	48	-56.00
4	12.4	32	-88.29
5	19.9	48	-82.77
6	92.5	128	-32.20
7	17.4	32	-59.11
8	13.3	32	-82.56
9	11.8	16	-30.22
10	17.9	32	-56.51
11	41.9	64	-41.74
12	38.5	64	-49.76
13	24.1	32	-28.16
14	48.7	80	-48.64
15	44.3	96	-73.70
16	23	32	-32.73
17	40.2	64	-45.68
18	66.4	96	-36.45
19	92.8	128	-31.88
20	51	80	-44.27
21	12.4	32	-88.29
22	55.7	80	-35.81
23	75.1	96	-24.43
24	38.3	64	-50.24
25	35.6	48	-29.67

Most variation between the two methods appears under 100 mg/kg of chlorides. This is because of the difference in sensitivities of the two methods, homogeneity of soils, and the different extraction factors (1:10 vs 1:4).