

Wildcat Measurement

Calibration Certificate

Pressure Recorder

Serial Number: 265-6213

Pressure Range 0-500# p.s.i. accuracy +/- 0.2 % Full Scale _____ p.s.i. _____

Increasing Pressure			Decreasing Pressure		
<i>Applied Pressure</i>	<i>Indicated Pressure</i>	<i>Difference</i>	<i>Applied Pressure</i>	<i>Indicated Pressure</i>	<i>Difference</i>
0.0#	0.0#	0.0#	400#	400#	0.0#
50#	50#	0.0#	300#	300#	0.0#
150#	150#	0.0#	200#	200#	0.0#
250#	250#	0.0#	100#	100#	0.0#
350#	350#	0.0#	0.0#	0.0#	0.0#
500#	500#	0.0#			

Calibrated By: DCT Gauge _____ Deadweight _____

This Is To Certify That This Recorder Has Been Inspected And Tested.

Remarks _____

Date Of Calibration 03/11/2002 Inspector 