

APPROVED

By Carl Chavez at 3:51 pm, Nov 02, 2020

Serial Number 901241
Model 5KPSIXP2I
Units PSI

Firmware Version R0324 Message Store -----

Run Index 2
Logging Type Actual

Logging Interval 90

Start Time 9/15/2020 13:36

Stop Time 9/15/2020 14:26

Time Reading

Event Event Data

Battery OK

Logging Interval, 90

Tare, -0.5

0	
0	
0	
0	582.6
90	581.6
180	581
270	580.2
360	579.5
450	579.3
540	579
630	578.4
720	578.1
810	578
900	577.7
990	577.3
1080	577
1170	576.8
1260	576.5
1350	576.6
1440	576.1
1530	575.9
1620	575.8
1710	575.5
1800	575.2
1890	575
1980	574.9
2070	574.6
2160	574.6
2250	574.2
2340	574.1
2430	573.9
2520	573.7
2610	573.5
2700	573.4
2790	573.2
2880	573
2970	573

Data Collection Report


Gauge Information

Serial Number	901241
Model	5KPSIXP2I
Message Store	-----
Units	PSI

Run Info

Start Time	9/15/20 1:36:20 PM
Stop Time	9/15/20 2:26:03 PM
Logging Interval	90

Pressure





9829 E. Easter Ave. • Centennial, CO 80112
303.794.8833 • Fax 303.730.1220
Toll Free 1.800.327.7257
www.jmcinstruments.com

CERTIFIED CALIBRATION

CUSTOMER Petrotek ORDER NO. _____

ITEM Digital Gauge RANGE 0-5000PSIG ITEM NO. 5035-2

TRUE VALUE	INDICATED VALUE	
	INCREASING READINGS	DECREASING READINGS
PSIG		
0.00	0	
500.00	499.5	499.6
1000.00	999.3	999.5
1500.00	1498.9	1498.9
2000.00	1999.2	1998.8
2500.00	2497.9	2497.5
3000.00	2997.7	2997.4
3500.00	3496.5	3495.7
4000.00	3995.8	3994.2
4500.00	4495.4	4494.0
5000.00	4995.3	4995.3

Tested On: Deadweight Tester S/N# 1GA4474

Traceable to National Institute of Standards and Technology certificate
17-043

Tested By: BMZ Date 17 January 2020

Remarks:

Crystal	XP2i	SN 901241
Accuracy is +/-	% of Full Scale or Better	
Test Conditions	70 °F; 612	mmHg Atm. Pressure

From: Dade, Randy <Lewis.Dade@HollyFrontier.com>
Sent: Monday, November 2, 2020 1:55 PM
To: Chavez, Carl J, EMNRD <CarlJ.Chavez@state.nm.us>
Cc: Dade, Randy <Lewis.Dade@HollyFrontier.com>
Subject: [EXT] RE: MIT Data for 2020

Carl,

I was present for the MIT's.

WDW-1: 30-015-27592: Witnessed by; Randy Dade, HFC; Nolan Beasley, Petrotek; Domingo Torres; Austin Services; Cory (Holt) Tilton, HFC
WDW-2: 30-015-20894: Witnessed by; Randy Dade, HFC; Wes Janes; Petrotek; Domingo Torres; Austin Services; Cory (Holt) Tilton, HFC
WDW-3: 30-015-26575: Witnessed by; Randy Dade, HFC; Nolan Beasley; Petrotek; Cory (Holt) Tilton, HFC
WDW-4: 30-015-44677: Witnessed By: Randy Dade, HFC; Nolan Beasley; Petrotek; Cory (Holt) Tilton, HFC

Lewis R. (Randy) Dade
Environmental Specialist
The HollyFrontier Companies
501 E. Main / P.O. Box 159
Artesia, NM 88210 / 88211-0159
575-746-5281 (o)
575-703-4735 (c)
575-746-5451 (f)
Email: Lewis.Dade@hollyfrontier.com