


D & L Meters & Instrument Service, Inc.

Lovington, NM 88260

P.O. Box 1621

Office: (575) 396-3715

Fax: (575) 396-5812



Date: Friday, March 27, 2015

30-015-32118

Certification of Pressure Recorder Test:

Company: Crane


Unit: NA

Model: Barton

Pressure Rating: 1,000#

Serial #: 12449

This Pressure Recorder was tested at midrange for accuracy and verified within +5% and -5% for 1,000# pressure element.



Issac Luna, Technician

