


HOBBS OCD

APR 10 2020

RECEIVED




**D & L Meters & Instrument Service, Inc.**

Lovington, NM 88260  
P.O. Box 1621  
Office: (575) 396-3715  
Fax: (575) 396-5812

**HOBBS OCD**

APR 10 2020

**RECEIVED**




**Date:** Tuesday, February 18, 2020

**Invoice #** \_\_\_\_\_

**Certification of Pressure Recorder Test:**

<b>Company:</b>	M & S
<b>Unit:</b>	M & S # 1
<b>Model:</b>	Barton
<b>Pressure Rating:</b>	1000#
<b>Serial #:</b>	N/A

***This Pressure Recorder was tested at midrange for accuracy and verified within +5% and -5% for a 1000# pressure element.***

  
\_\_\_\_\_  
Jesse Arenivas, Technician