



GRAPHIC CONTROLS CORPORATION
BUFFALO, NEW YORK

LBS.
PER
SQ. IN.

API 30-025-33547
Sble S #1 BKI-28

BP 4544

Key Energy Co.
Formation Test

Start 10:27 AM
9/10/10

STOP
11:29
9/10/10

HE-REP-3
Witnessed by J. Grisham
V. Grisham

MIDNIGHT

NOON

6 PM

6 AM

American Valve & Meter, Inc.

1113 W. BROADWAY

P.O. BOX 166

HOBBS, NM 88240

TO: Key Energy

DATE: 09-08-10

This is to certify that:

I, Bud Collins, Technician for American Valve & Meter,

Inc., has checked the calibration of the following instrument.

8" Pressure recorder Serial No: 8131

at these points.

Pressure 0 - 500 ~~75~~ Temperature _____

<u>Test</u>	<u>Found</u>	<u>Left</u>	<u>Test</u>	<u>Found</u>	<u>Left</u>
<u>0</u>	—	<u>0</u>	—	—	—
<u>250</u>	—	<u>250</u>	—	—	—
<u>500</u>	—	<u>500</u>	—	—	—
<u>350</u>	—	<u>350</u>	—	—	—
<u>100</u>	—	<u>100</u>	—	—	—
<u>0</u>	—	<u>0</u>	—	—	—

Remarks: _____

Signature Bud Collins

