

NEW-TEX LAB
P.O. Box 1161
Hobbs, New Mexico 88240

ANALYSIS CERTIFICATE

CLIENT: JOHN WEST ENGINEERING
ADDRESS: 412 N. DAL PASO
CITY, STATE: HOBBS, NM 88240

ANALYSIS NUMBER: 1885
DATE OF RUN: 013086
DATE SECURED: 012986

SAMPLE IDENT: MARSHALL PIPE & SUPPLY - SPEIGHT NO.1
SAMPLING PRESS: SAMPLING TEMP:

SAMPLE FROM TUBING

***** GAS ANALYSIS *****

	MOLE PERCENT	GAL/ MCF
NITROGEN	9.369	
CARBON DIOXIDE	0.336	
METHANE	76.928	
ETHANE	6.105	1.628
PROPANE	4.330	1.189
ISO-BUTANE	0.606	0.198
NORMAL BUTANE	1.347	0.423
ISO-PENTANE	0.337	0.124
NORMAL PENTANE	0.322	0.116
HEXANES	0.300	0.123
TOTAL	100.000	3.801

PROPANE GPM: 1.19 BUTANES GPM: 0.62
ETHANE GPM: 1.63 PENTANES PLUS GPM: 0.36

SPECIFIC GRAV (CALC): 0.7163
MOLE WEIGHT: 20.75

PSV-370/01 FT	PRESSURE (PSIA)	WET	DRY
	14.676	1079	1079
	14.650	1076	1095
	14.720	1082	1101
	14.735	1082	1102

ANALYZED BY:

Deane Simpson