

American Valve & Meter, Inc.

1113 W. Broadway
P.O. Box 166
Hobbs, NM 88240

To: A+B TRUCKING

Date: 2-7-02

This is to Certify that:

I, Tony Flores, Technician for American Valve & Meter, Inc.,

has checked the calibration of the following instrument

8" Single pen P.S.I. recorder

Serial Number None

at these points.


Pressure 0-1000#

Temperature _____

Test	Found	Left
<u>0</u>	<u>+ 0</u>	<u>0</u>
<u>200</u>	<u>400</u>	<u>200</u>
<u>400</u>	<u>600</u>	<u>400</u>
<u>600</u>	<u>800</u>	<u>600</u>
<u>800</u>	<u>1000</u>	<u>800</u>
<u>1000</u>	<u>off chart</u>	<u>1000</u>

Test	Found	Left
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

Remarks: _____



Signature Tony Flores