

METERING & TESTING SERVICES, INC.
CALIBRATION REPORT

COMPANY: <i>METERING & TESTING</i>	LEASE: <i>Shop</i>	DATE: <i>3-29-2002</i>
COUNTY: <i>MIDLAND</i>	STATE: <i>TX.</i>	LOCATION: <i>Shop</i>
PURCHASER:	PIPELINE CONNECTION:	STATION NUMBER:
MAKE OF METER: <i>METSEACO</i>	SERIAL NUMBER: <i>MFG 0521</i>	GAS GRAVITY:
DIFF. RANGE:	STATIC RANGE: <i>1500 X 1500</i>	TEMP. RANGE:
AVERAGE DIFF:	AVERAGE STATIC:	AVERAGE TEMP:
LINE SIZE:	UPSTREAM:	DOWNSTREAM:
ORIFICE SIZE:	ORIFICE CONDITION:	SEAL CONDITION:
FLANGE OR PIPE TAP:	VANES:	CALC. BETA RATIO:
PEN ARC: <i>O.K.</i>	PEN DRAG: <i>O.K.</i>	CLOCK ROTATION: <i>24 Hr.</i>

METER CALIBRATION DATA

1500 X 1500

DIFFERENTIAL STATIC				STATIC				TEMPERATURE			
FOUND		LEFT		FOUND		LEFT		FOUND		LEFT	
PK	DIFF	PK	DIFF	DW	STATIC	DW	STATIC	THEM. TEMP.		THEM. TEMP.	
0-0		0-0		0-0		0-0					
300-300		100-100		300-300		100-100					
600-600		300-300		600-600		300-300					
900-900		600-600		900-900		600-600					
1200-1200		900-900		1200-1200		900-900					
1500-1500		1200-1200		1500-1500		1200-1200					
		1500-1500				1500-1500					
		1200-1205				1200-1205					
		900-905				900-905					
		600-605				600-605					
		300-305				300-305					
		100-100				100-100					
		0-0				0-0					

REMARKS: METER (WAS) / WAS NOT IN CALIBRATION AS FOUND.

TESTER: *B. Elliott*

WITNESS:

WITNESS:

RECEIVED

MAY 29 2002

Environmental Bureau
Oil Conservation Division